

ddPCR™ Copy Number Variation Assays, Validated

Catalog #	Description
10031240	ddPCR CNV Assay, Validated (FAM) , 200 x 20 µl reactions
10031241	ddPCR CNV Assay, Validated (FAM) , 1,000 x 20 µl reactions
10031242	ddPCR CNV Assay, Validated (FAM) , 2,500 x 20 µl reactions
10031243	ddPCR CNV Assay, Validated (HEX) , 200 x 20 µl reactions
10031244	ddPCR CNV Assay, Validated (HEX) , 1,000 x 20 µl reactions
10031245	ddPCR CNV Assay, Validated (HEX) , 2,500 x 20 µl reactions

For research purposes only.

Description

ddPCR Copy Number Variation (CNV) Assays, Validated, are expertly designed specifically for Droplet Digital™ PCR (ddPCR) using proprietary computational algorithms. These assays have been wet-lab validated by Bio-Rad™ Laboratories, Inc. Validation data are available for review and download from the assay page on the Bio-Rad website. Reference assays with the complementary fluorophores can be purchased separately.

Ordering Information

Visit bio-rad.com/digital-assays to order ddPCR CNV Assays.

Storage and Stability

ddPCR CNV Assays are stable for 12 months when stored at 4°C protected from light. The 20x assay mix can be kept at -20°C for long-term storage.

Kit Contents

The ddPCR CNV Assay is a 20x concentrated, ready-to-use primer-probe mix. Each kit comes with 200, 1,000, or 2,500 µl of the 20x assay mix (18 µM primers and 5 µM probe), sufficient for 200, 1,000, or 2,500 x 20 µl reactions, respectively.

Reagents and Equipment

For assays using the QX200™ Droplet Generator (catalog #1864002) or Automated Droplet Generator (#1864101):

- For 1–2 targets, ddPCR Supermix for Probes (No dUTP) (#1863023, #1863024, #1863025) is recommended
- For >2 targets, ddPCR Multiplex Supermix (#12005909, #12005910, #12005911) is recommended
- QX200 Droplet Reader (#1864003) or QX600™ Droplet Reader (#12013328)
- PX1 PCR Plate Sealer (#1814000)

For assays using the QX ONE™ Droplet Digital PCR System (#12006536):

- ddPCR Multiplex Supermix (#12005909, #12005910, #12005911)
- PX1 PCR Plate Sealer (#1814000)

Refer to the QX200 Droplet Generator Instruction Manual (10031907), QX ONE Droplet Digital PCR System and QX ONE Software User Guide (10000116512), or Automated Droplet Generator Instruction Manual (10043138) for ordering information on consumables such as oils, cartridges, gaskets, plates, and seals. Refer to Table 1 for fluorophore compatibility with instruments.

Table 1. Fluorophore compatibility.

QX200 Droplet Reader	QX ONE ddPCR System	QX600 Droplet Reader
FAM	FAM	FAM
HEX	HEX	HEX

Restriction Enzyme Digestion of Sample DNA

DNA fragmentation by restriction digestion prior to droplet generation enables optimal accuracy by separating tandem gene copies, reducing sample viscosity, and improving template accessibility. Two strategies may be used to perform restriction digestion of DNA samples: digestion directly in the ddPCR reaction during setup (recommended) or conventional digestion prior to Droplet Digital PCR.

Digestion Directly in ddPCR Reaction

- Efficient digestion of sample DNA can be achieved by direct addition of restriction enzyme to the ddPCR reaction
- 2–5 units of restriction enzyme per ddPCR reaction are recommended
- Dilute the restriction enzyme using the recommended diluent buffer according to the manufacturer's instructions, and then add 1 µl to the ddPCR reaction according to the guidelines in Table 2
- Reactions can be set up at room temperature; no additional incubation time is required
- The addition of restriction enzyme buffers with high salt content can inhibit Droplet Digital PCR and should be avoided

Digestion Prior to Droplet Digital PCR

- Restriction enzyme digestion can be carried out as a separate reaction before ddPCR reaction setup
- Use 10–20 units of restriction enzyme per microgram of genomic DNA (gDNA)
- Incubate the reaction for 1 hr at the temperature recommended for the restriction enzyme
- Heat inactivation is not required but can be considered if long-term storage is required; do not heat inactivate above 65°C
- DNA purification is not necessary after restriction digestion
- Use a minimum 10-fold dilution of the digest to reduce the salt content of the sample in Droplet Digital PCR
- Store digested DNA at –20°C or below

Reaction Setup

- For most routine CNV applications where a diploid target copy number is expected to be 10 or less, ≤50 ng of human gDNA should be added per ddPCR well. If 10–50 copies per diploid genome are expected in a sample, add ≤15 ng of sample per well. For copy number evaluation >50 copies per diploid genome, strategies using multiple wells can be used.
- Thaw all components to room temperature. Mix thoroughly by vortexing each tube to ensure homogeneity because a concentration gradient may form during –20°C storage. Centrifuge briefly to collect contents at the bottom of each tube and store protected from light.
- Prepare samples at the desired concentration before setting up the reaction mix according to the guidelines in Table 2. If multiple samples are to be assayed using the same target and reference duplex, prepare a master reaction mix without sample template, dispense equal aliquots into the reaction tubes, and add the sample template to each reaction tube as the final step.

Table 2. Preparation of the reaction mix.

Component	Volume per Reaction, μ l	Final Concentration
2x ddPCR Supermix for Probes (No dUTP)	10	1x
20x target primers/probe (FAM)	1	1x***
20x reference primers/probe (HEX)	1	1x***
Restriction enzyme, diluted*	1	2–5 U/reaction
DNA sample or water	Variable	≤50 ng†
Total volume	20**	—

* This component should be replaced by water if digestion is performed prior to Droplet Digital PCR.

** For the Automated Droplet Generator, prepare 22 μ l per well.

*** 900 nM primers/250 nM each probe.

† For most human gDNA experiments. See loading recommendations in step 1 of this section.

- Mix thoroughly by vortexing each tube. Centrifuge briefly to ensure that all components are at the bottom of each reaction tube. Allow reaction tubes to equilibrate at room temperature for about 3 minutes.
- Transfer the reaction mix from the reaction tubes to the appropriate ddPCR Cartridge as follows.
 - For the QX200 Droplet Generator, load 20 μ l of each reaction mix into a sample well of a DG8 Cartridge. Follow subsequent instructions as specified in the QX200 Droplet Generator Instruction Manual (10031907)
 - For the Automated Droplet Generator, follow instructions in the Automated Droplet Generator Instruction Manual (10043138)
 - For the QX ONE ddPCR System, load 20 μ l of each reaction mix into a sample well of a GCR96 Cartridge. Follow subsequent instructions as specified in the QX ONE Droplet Digital PCR System and QX ONE Software User Guide (10000116512)

Thermal Cycling Conditions

Follow instructions for thermal cycling based on the droplet generator used.

- For the QX200 Droplet Generator, carefully transfer droplets into a clean 96-well plate. Seal the plate using the PX1 PCR Plate Sealer at 180°C for 5 sec. Proceed to thermal cycling (see Table 3)
- For the Automated Droplet Generator, remove the plate containing ddPCR droplets from the Automated Droplet Generator. Seal the plate using the PX1 PCR Plate Sealer at 180°C for 5 sec. Proceed to thermal cycling (see Table 3)
- For the QX ONE ddPCR System, thermal cycling is integrated into and sequentially performed by the system itself. Hence, no additional equipment or sample handling is required for this step. Refer to the QX ONE Droplet Digital PCR System and QX ONE Software User Guide (10000116512) for plate setup instructions. Use appropriate thermal cycling conditions as specified in Table 3

Table 3. Thermal cycling conditions.*

Cycling Step	Temperature, °C	Time	Number of Cycles	
Hold (QX ONE ddPCR System only)	25	3 min	1	
Enzyme activation	95	10 min	1	
Denaturation	94	30 sec	40	
Annealing/extension	60	1 min**	40	
Enzyme deactivation	98	10 min	1	
Hold	QX200 or QX600 ddPCR System (optional)	4	Infinite	1
	QX ONE ddPCR System (required)	25	1 min	1

* For the C1000 Touch Thermal Cycler, use a heated lid set to 105°C and set the sample volume to 40 μ l.

** Check/adjust ramp rate settings to ~2°C/sec.

Data Acquisition and Analysis

Follow instructions for data acquisition and analysis based on the droplet reader in use.

- For the QX200 Droplet Reader, refer to the QX200 Droplet Reader and QX Manager Software Standard Edition User Guide (10000107223) or the QX200 Droplet Reader and QX Manager Software Regulatory Edition User Guide (10000107224)
- For the QX600 Droplet Reader, refer to the QX600 Droplet Reader and QX Manager Software Standard Edition User Guide (10000153877) or the QX600 Droplet Reader and QX Manager Software Premium Edition User Guide (10000153878)
- For the QX ONE ddPCR System, refer to the QX ONE Droplet Digital PCR System and QX ONE Software User Guide (10000116512) and the QX ONE Software User Guide for Standard Edition (10000116655) or Regulatory Edition (10000116656)



BIO-RAD, DDPCR, DROPLET DIGITAL, QX200, QX600, and QX ONE are trademarks of Bio-Rad Laboratories, Inc. in certain jurisdictions. All trademarks used herein are the property of their respective owner. © 2022 Bio-Rad Laboratories, Inc.

Purchase of Digital PCR and/or Single-Cell NGS Sample Preparation products (the "Products") from Bio-Rad Laboratories is subject to Bio-Rad Laboratories, Inc. Standard Terms and Conditions of Sale, which can be accessed at <https://www.bio-rad.com/en-us/terms-conditions>. Unless we expressly state otherwise in additional Terms and Conditions, no rights are granted for you to distribute or resell the Products. Unless we expressly state otherwise in additional Terms and Conditions, no rights are granted for the development or commercialization of diagnostic assays for use with the Products without a license from Bio-Rad. The Products and/or their use are covered by U.S. and foreign patents and/or pending patent applications owned by or under license to Bio-Rad Laboratories, Inc. See <https://www.bio-rad.com/en-us/trademarks>.